

TELE-TRAFFIC (UK) LIMITED

LaserTec Centre
C2 Harris Road
WARWICK
CV34 5JU

Tel: 01926-407272
Fax: 01926-407977
Email: tt @ teletraffic . co. uk

CERTIFICATE OF CONFORMITY

Certificate Number: **18 -0373**

Date: **13 April 2018**

Customer: **DYFED-POWYS POLICE**

Instrument : **LTI 20-20 UltraLyte 1000**

Manufacturer: **LASER Technology inc**

Serial Number: **UX 022186**

U.K. Serial Number: **449012UX**

This instrument has been calibrated/calibration verified and complies with the manufacturers specification at the measured points. Calibration/verification was controlled by the use of documented procedures using equipment traceable to National standards.

Approved Signatory



for Tele-Traffic (UK) Ltd.

TELE-TRAFFIC (UK) LIMITED

CALIBRATION DATA SHEET

LTI 20-20 UltraLyte 1000

Certificate No 18-0373

MANUFACTURED BY: LASER TECHNOLOGY INC.

MANUFACTURERS SERIAL NUMBER: UX 022186

UNITED KINGDOM SERIAL NUMBER: 449012UX

Visual Inspection: Pass
Initialisation Test: Pass
Display Test: Pass
Serial Port Check: Pass

Distance Checks

| | | | |
|------|------------|------|--------------|
| No:1 | 25 Metres | Pass | Speed: 0 mph |
| No:2 | 75 Metres | Pass | Speed: 0 mph |
| No:3 | 100 Metres | Pass | Speed: 0 mph |

Pulse Repetition Frequency Test: 125 pps nominal
Output Period Specification: 7.9968 to 8.00 ms
Measured Output Period: 8.000 ms

Speed Simulation Tests

Specification: +/- 1MPH or 2 KPH

| | | |
|--------------------|---------|------|
| Measured Test No:1 | @ 20MPH | = 20 |
| Measured Test No 2 | @ 30MPH | = 30 |
| Measured Test No 3 | @ 50MPH | = 50 |
| Measured Test No:4 | @ 70MPH | = 70 |
| Measured Test No:5 | @ 80MPH | = 79 |

Scope Alignment Test: Pass
(Horizontal and Vertical alignment correct)

NOTES

Any parts or components that may have been replaced during the calibration procedure are certified as identical to those used in the device as type approved.

I CERTIFY THAT THE FOREGOING INTERNAL TESTS AND TEST DATA ARE CORRECT AND THAT THE INSTRUMENT ALIGNMENT IS CORRECT

TESTED BY: James Sample

DATE: : 13 April 2018
CAL DUE: : 13 April 2019

TIME: 10:42 AM