

TELE-TRAFFIC (UK) LIMITED

CALIBRATION DATA SHEET

LTI 20-20 TruCAM

Certificate No 19-0016

MANUFACTURED BY: LASER TECHNOLOGY INC.

MANUFACTURERS SERIAL NUMBER: TC 003635

UNITED KINGDOM SERIAL NUMBER: 449510TC

Visual Inspection: Pass

Initialisation Test: Pass

Display Test: Pass

Firmware Version Check: Pass

Distance Checks

No:1 25 Metres
No:2 75 Metres
No:3 100 Metres

Pass
Pass
Pass

Speed: 0 mph
Speed: 0 mph
Speed: 0 mph

Pulse Repetition Frequency Test

Output Period Specification:

Measured Output Period:

200.321 ppm nominal
200.221 to 200.421 Hz
200.319 Hz

Speed Simulation Tests

Specification:

Measured Test No:1
Measured Test No:2
Measured Test No:3
Measured Test No:4
Measured Test No:5

+/- 1MPH or 2 KPH

@ 20MPH = 19
@ 30MPH = 30
@ 50MPH = 49
@ 70MPH = 69
@ 80MPH = 80

Scope Alignment Test:

(Horizontal and Vertical alignment correct)

Pass

NOTES

Any parts or components that may have been replaced during the calibration procedure are certified as identical to those used in the device as type approved.

I CERTIFY THAT THE FOREGOING INTERNAL TESTS AND TEST DATA ARE CORRECT AND THAT THE INSTRUMENT ALIGNMENT IS CORRECT

TESTED BY:

James Sample

DATE: 07 January 2019
CAL DUE: 07 January 2020

TIME: 12:23 PM

TELE-TRAFFIC (UK) LIMITED

LaserTec Centre
C2 Harris Road
WARWICK
CV34 5JU

Tel: 01926-407272
Fax: 01926-407977
Email: tt @ teletraffic . co. uk

CERTIFICATE OF CONFORMITY

Certificate Number: **19 -0016**

Date: **07 January 2019**

Customer: **NORTH WALES POLICE**

Instrument : **LTI 20-20 TruCAM**

Manufacturer: **LASER Technology inc**

Serial Number: **TC 003635**

U.K. Serial Number: **449510TC**

This instrument has been calibrated/calibration verified and complies with the manufacturers specification at the measured points. Calibration/verification was controlled by the use of documented procedures using equipment traceable to National standards.

Approved Signatory



for Tele-Traffic (UK) Ltd.