

CALIBRATION CERTIFICATE

for
Kustom LASERwitness Lite
Mobile Speed Enforcement System

Customer: Gwent Police **Serial Number:** LW06289
Certificate Number: T5501
Date of Calibration: 23 August 2018 **Next Calibration:** 22 August 2019
Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 20.000178 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming Vertical - Pass
 Horizontal - Pass
Eye Safety - Pass

Range \pm 0.5 ft

Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.2 ft	484.3 ft	488.5 ft

Speed \pm 1 MPH

Actual	22 MPH	-34 MPH	48 MPH	-52 MPH	67 MPH
Indicated	22 MPH	-34 MPH	48 MPH	-52 MPH	67 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK



L. Korcz

Technical Signatory

TRU-CAL-CER-005

TRUVELO (UK) LTD

Tel: 020 8847 4400

email: info@truelouk.com