

CALIBRATION CERTIFICATE

for
Kustom LASERwitness Lite
Mobile Speed Enforcement System

Customer: South Wales Police **Serial Number:** LW06275

Certificate Number: T5482

Date of Calibration: 08 August 2018 **Next Calibration:** 07 August 2019

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 20.000045 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming Vertical - Pass
 Horizontal - Pass
Eye Safety - Pass

Range ± 0,5 ft

Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.2 ft	284.3 ft	488.4 ft

Speed ± 1 MPH

Actual	71 MPH	55 MPH	28 MPH	-74 MPH	-30 MPH
Indicated	71 MPH	55 MPH	28 MPH	-74 MPH	-30 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK



L. Kórcz
Technical Signatory

TRUVELO (UK) LTD
