CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer:

South Wales Police

Serial Number:

LW06290

Certificate Number:

T5493

Date of Calibration:

21 August 2018

Next Calibration:

20 August 2019

Place of Calibration:

14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum

- CC85

Oscillator Frequency

- 20.000077 MHz

Self Test Sequence

- Pass

External Controls

- Pass

- Pass

Aiming

- Pass

Eye Safety

- Pass

Range ± 0.5 ft

Actual

156.3 ft

284.5 ft

488.4 ft

Indicated 156.2 ft

Vertical

Horizontal

484.3 ft

488.5 ft

Speed ± 1 MPH

Actual

21 MPH

-34 MPH

-47 MPH

51 MPH

68 MPH

Indicated

21 MPH

-34 MPH

-47 MPH

51 MPH

68 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

L Korcz

Technical Signatory

TRU-CAL-CER-005

TRUVELO (UK) LTD

Tel:020 8847 4400

email:info@truvelouk.com