

# TELE-TRAFFIC (UK) LIMITED

LaserTec Centre  
C2 Harris Road  
WARWICK  
CV34 5JU

Tel: 01926-407272  
Fax: 01926-407977  
Email: tt @ teletraffic . co. uk

## TELE-TRAFFIC CERTIFICATE OF CONFORMITY

Certificate Number: 18 -0137

Date: 08 February 2018

Customer: DYFED-POWYS POLICE

Instrument : LTI 20-20 TruCAM

Manufacturer: LASER Technology inc


Serial Number: TC 003892

U.K. Serial Number: 449529TC

This instrument has been calibrated/calibration verified and complies with the manufacturers specification at the measured points. Calibration/verification was controlled by the use of documented procedures using equipment traceable to National standards.

## TELE-TRAFFIC

Approved Signatory



for Tele-Traffic (UK) Ltd.

# TELE-TRAFFIC (UK) LIMITED CALIBRATION DATA SHEET

LTI 20-20 TruCAM

Certificate No 18- 0137

MANUFACTURED BY: LASER TECHNOLOGY INC.

MANUFACTURERS SERIAL NUMBER: TC 003892

UNITED KINGDOM SERIAL NUMBER: 449529TC

Visual Inspection: Pass  
Initialisation Test: Pass  
Display Test: Pass  
Serial Port Check: Pass

#### Distance Checks

No:1	25 Metres	Pass	Speed: 0 mph
No:2	75 Metres	Pass	Speed: 0 mph
No:3	100 Metres	Pass	Speed: 0 mph

#### Pulse Repetition Frequency Test

200.321 ppm nominal  
Output Period Specification: 200.221 to 200.421 Hz  
Measured Output Period: 200.320 Hz

#### Speed Simulation Tests

Specification: +/- 1MPH or 2 KPH

Measured Test No:1	@ 20MPH	= 19
Measured Test No 2	@ 30MPH	= 30
Measured Test No 3	@ 50MPH	= 49
Measured Test No:4	@ 70MPH	= 69
Measured Test No:5	@ 80MPH	= 80

#### Scope Alignment Test:

(Horizontal and Vertical alignment correct)

Pass

Any parts or components that may have been replaced during the calibration procedure are certified as identical to those used in the device as type approved.

I CERTIFY THAT THE FOREGOING INTERNAL TESTS AND TEST DATA ARE CORRECT  
AND THAT THE INSTRUMENT ALIGNMENT IS CORRECT

TESTED BY: James Sample

DATE: : 08 February 2018  
CAL DUE: : 08 February 2019

TIME: 10:45 AM