CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer:	Gwent Police	**Serial Number:**	LW06264
		Certificate Number:	T5583
Date of Calibration:	16 October 2018	**Next Calibration:**	15 October 2019

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum		- CC85
Oscillator Frequency		- 19.999867 MHz
Self Test Sequence		- Pass
External Controls		- Pass
Aiming	Vertical	- Pass
	Horizontal	- Pass
Eye Safety		- Pass

Range ± 0.5 ft					
Actual	156.3 ft	284.5 ft	488.4 ft		
Indicated	156.3 ft	284.4 ft	488.4 ft		
Speed ± 1 MPH					
Actual	72 MPH	-70 MPH	57 MPH	-23 MPH	31 MPH
Indicated	72 MPH	-70 MPH	57 MPH	-23 MPH	31 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

L Korcz
Technical Signatory

TRU-CAL-CER-005

TRUVELO (UK) LTD

Tel:020 8847 4400 email:info@truvelouk.com