

## TELE-TRAFFIC (UK) LIMITED

LaserTec Centre  
C2 Harris Road  
WARWICK  
CV34 5JU

Tel: 01926-407272  
Fax: 01926-407977  
Email: tt @ teletraffic . co. uk

### CERTIFICATE OF CONFORMITY

Certificate Number: 19 -0135

Date: 06 February 2019

Customer: DYFED-POWYS POLICE

Instrument : LTI 20-20 TruCAM

Manufacturer: LASER Technology inc

Serial Number: TC 004152

U.K. Serial Number: 449572TC

This instrument has been calibrated/calibration verified and complies with the manufacturers specification at the measured points. Calibration/verification was controlled by the use of documented procedures using equipment traceable to National standards.

Approved Signatory



for Tele-Traffic (UK) Ltd.

# TELE-TRAFFIC (UK) LIMITED

## CALIBRATION DATA SHEET

LTI 20-20 TruCAM

Certificate No 19- 0135

MANUFACTURED BY: LASER TECHNOLOGY INC.

MANUFACTURERS SERIAL NUMBER: TC 004152

UNITED KINGDOM SERIAL NUMBER: 449572TC

Visual Inspection: Pass  
Initialisation Test: Pass  
Display Test: Pass  
Firmware Version Check: Pass

#### Distance Checks

No:1	25 Metres	Pass	Speed: 0 mph
No:2	75 Metres	Pass	Speed: 0 mph
No:3	100 Metres	Pass	Speed: 0 mph

Pulse Repetition Frequency Test 200.321 ppm nominal  
Output Period Specification: 200.221 to 200.421 Hz  
Measured Output Period: 200.319 Hz

#### Speed Simulation Tests

Specification: +/- 1MPH or 2 KPH

Measured Test No:1	@ 20MPH	= 20
Measured Test No 2	@ 30MPH	= 29
Measured Test No 3	@ 50MPH	= 50
Measured Test No:4	@ 70MPH	= 69
Measured Test No: 5	@ 80MPH	= 80

Scope Alignment Test: Pass  
(Horizontal and Vertical alignment correct)

#### NOTES

Any parts or components that may have been replaced during the calibration procedure are certified as identical to those used in the device as type approved.

I CERTIFY THAT THE FOREGOING INTERNAL TESTS AND TEST DATA ARE CORRECT  
AND THAT THE INSTRUMENT ALIGNMENT IS CORRECT

TESTED BY: James Sample DATE: : 06 February 2019 TIME: 3:30 PM  
CAL DUE: : 06 February 2020