

TELE-TRAFFIC (UK) LIMITED

LaserTec Centre
C2 Harris Road
WARWICK
CV34 5JU

Tel: 01926-407272
Fax: 01926-407977
Email: tt @ teletraffic . co. uk

CERTIFICATE OF CONFORMITY

Certificate Number: **19 -0515**

Date: **21 May 2019**

Customer: **SOUTH WALES POLICE**

Instrument : **LTI 20-20 TruCAM**

Manufacturer: **LASER Technology inc**

Serial Number: **TC 003652**

U.K. Serial Number: **449527TC**

This instrument has been calibrated/calibration verified and complies with the manufacturers specification at the measured points. Calibration/verification was controlled by the use of documented procedures using equipment traceable to National standards.

Approved Signatory



for Tele-Traffic (UK) Ltd.

TELE-TRAFFIC (UK) LIMITED

CALIBRATION DATA SHEET

LTI 20-20 TruCAM

Certificate No 19- 0515

MANUFACTURED BY: LASER TECHNOLOGY INC.

MANUFACTURERS SERIAL NUMBER: TC 003652

UNITED KINGDOM SERIAL NUMBER: 449527TC

Visual Inspection:	Pass
Initialisation Test:	Pass
Display Test:	Pass
Firmware Version Check:	Pass

Distance Checks

No:1	25 Metres	Pass	Speed: 0 mph
No:2	75 Metres	Pass	Speed: 0 mph
No:3	100 Metres	Pass	Speed: 0 mph

Pulse Repetition Frequency Test:	200.321 ppm nominal
Output Period Specification:	200.221 to 200.421 Hz
Measured Output Period:	200.320 Hz

Speed Simulation Tests

Specification:	+/- 1MPH or 2 KPH	
Measured Test No:1	@ 20MPH	= 20
Measured Test No 2	@ 30MPH	= 30
Measured Test No 3	@ 50MPH	= 50
Measured Test No:4	@ 70MPH	= 70
Measured Test No:5	@ 80MPH	= 79

Scope Alignment Test:

(Horizontal and Vertical alignment correct)

Pass

NOTES

Any parts or components that may have been replaced during the calibration procedure are certified as identical to those used in the device as type approved.

I CERTIFY THAT THE FOREGOING INTERNAL TESTS AND TEST DATA ARE CORRECT AND THAT THE INSTRUMENT ALIGNMENT IS CORRECT

TESTED BY: James Sample

DATE: : 21 May 2019
CAL DUE: : 21 May 2020

TIME: 12:02 PM