

CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite Mobile Speed Enforcement System

Customer: Gwent Police **Serial Number:** LW06267

Certificate Number: 2956

Date of Calibration: 20 June 2019 **Next Calibration:** 19 June 2020

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 19.999822 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
Horizontal - Pass
Eye Safety - Pass

Range ± 0.5 ft

Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.2 ft	284.3 ft	488.3 ft

Speed ± 1 MPH

Actual	74 MPH	-70 MPH	59 MPH	-25 MPH	21 MPH
Indicated	74 MPH	-70 MPH	59 MPH	-25 MPH	21 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK


L Korca

Technical Signatory

TRUVELO (UK) LTD

TRU-CAL-CER-005

Tel: 020 8847 4400

email: info@truelouk.com