

# CALIBRATION CERTIFICATE

for  
**Kustom LASERwitness Lite**  
Mobile Speed Enforcement System

**Customer:** South Wales Police **Serial Number:** LW06282

**Certificate Number:** 3105

**Date of Calibration:** 08 August 2019 **Next Calibration:** 07 August 2020

**Place of Calibration:** 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

## Calibration Results:

Checksum - CC85  
Oscillator Frequency - 20.000038 MHz  
Self Test Sequence - Pass  
External Controls - Pass  
Aiming - Vertical - Pass  
                    - Horizontal - Pass  
Eye Safety - Pass

## Range ± 0.5 ft

Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.2 ft	284.5 ft	488.6 ft

## Speed ± 1 MPH

Actual	53 MPH	-27 MPH	-41 MPH	20 MPH	71 MPH
Indicated	53 MPH	-27 MPH	-41 MPH	20 MPH	71 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

  
L Korcz  
Technical Signatory

---

# TRUVELO (UK) LTD

---